

CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

DATE OF ISSUE: 12 December 2023 CERTIFICATE NUMBER: **1154279**



0654



Unit 5 Octavian Way
Team Valley Trading Estate
Gateshead
NE11 0HZ

Tel: +44 (0)191 4875951

Page 1 of 3

Approved Signatory

Electronically Authorised Document

P K CLARK J FRYER
 R J WADE M FOY
 M A FROST
 M S PARDOE

CUSTOMER

DJB LABCARE LTD
UNIT 12 HOWARD WAY
CROMWELL BUSINESS CENTRE
NEWPORT PAGNELL
BUCKINGHAMSHIRE
MK16 9QS
UNITED KINGDOM

MANUFACTURER

DESCRIPTION

MODEL

SERIAL No.

IDENT No.

DATE RECEIVED

DATE OF CALIBRATION

ORDER No.

STANDARD

TACHOMETER OPTICAL

AT-6

140422874

UNKNOWN

7 DECEMBER 2023

11 DECEMBER 2023

NS23605

INSTRUMENT CONDITION

Adjustments Made
Repairs Made

No
No

ENVIRONMENT

The instrument was placed in the Laboratory environment for a minimum period of 4 hours prior to calibration.

The ambient conditions were: 18 to 23°C and 45 %rh \pm 15 %rh.

STABILITY

The results contained in this Certificate refer to the measurements made at the time of test and not to the instrument's ability to maintain calibration.

PROCEDURE

Measurements were performed in accordance with the in house Laboratory procedure No.1053

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654



CERTIFICATE NUMBER

1154279

Page 2 of 3

Calibration Equipment Used:

| Cert Number | Ident Number | Model | Serial Number | Test Equipment Calibration Due |
|--------------------|---------------------|--------------|----------------------|---|
| 1150572IH | 143 | 33220A | MY44025316 | 28 Sep 2024 |
| 1149140IH | 475 | A002206.G1 | 13572091 | 4 Sep 2024 |

Notes:

Results relate only to the items calibrated.

CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654



CERTIFICATE NUMBER

1154279

RESULT SHEET 1053-NC-UKAS : STANDARD AT-6 OPTICAL TACHOMETER

AS FOUND

BATTERY REPLACED: YES

The Revolution Per Minute (RPM) was simulated using a precision Frequency Generator with an attached optical light source.

RPM

| Range | Simulated | Indicated | Units | Indicated Deviation | Measurement Uncertainty \pm |
|------------------|-----------|-----------|-------|---------------------|-------------------------------|
| 2 to 99999RPM | 600.000 | 600.0 | RPM | 0.000 | 0.064 |
| | 1000.00 | 1000 | RPM | 0.00 | 0.58 |
| | 2500.00 | 2500 | RPM | 0.00 | 0.58 |
| | 5000.00 | 5000 | RPM | 0.00 | 0.58 |
| | 7500.00 | 7500 | RPM | 0.00 | 0.58 |
| | 9999.86 | 10000 | RPM | 0.14 | 0.58 |
| | 14999.52 | 15000 | RPM | 0.48 | 0.58 |
| | 19999.26 | 20000 | RPM | 0.74 | 0.58 |
| | 24998.91 | 25000 | RPM | 1.09 | 0.58 |
| | 29998.47 | 30000 | RPM | 1.53 | 0.58 |

Manufacturer stated Accuracy: $\pm(0.05\%+1 \text{ digit})$

COMMENTS:

TEST ENGINEER: M Foy

DATE: 11th December 2023