



## Certificate of Calibration

Certificate number: 14/020222

Equipment: Fastime 3 Stopwatch  
Serial number: 014

Date: 02/02/2022  
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

### Results:

Set Value	Measured Value
05:00.00	05:00.04
10:00.00	10:00.09
60:00.00	60:00.03

Method of test:  
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam

