



12 Howard Way  
Cromwell Business Centre  
Newport Pagnell  
Bucks MK16 9QS

Tel: 01908 612598  
Fax: 01908 217974

E-mail: [service@djblabcare.co.uk](mailto:service@djblabcare.co.uk)  
Website: [www.djblabcare.co.uk](http://www.djblabcare.co.uk)

## Certificate of Calibration

Certificate number: 18-240822

Equipment: Fastime 3 Stopwatch  
Serial number: 018

Date: 24/08/2022  
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

### Results:

| Set Value | Measured Value |
|-----------|----------------|
| 05:00.00  | 04:59.96       |
| 10:00.00  | 09:59.94       |
| 60:00.00  | 60:00.04       |

### Method of test:

All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam